



PT Calendar JAN. 26 - DEC. 26

Octagon Measurement Solutions Pvt. Ltd.

Lab Address :S.No.15 A/2, J7-J8, GKD Industrial Estate, Nandedgaon,
Sinhgad Road, Pune- 411041, Maharashtra.

SN	Category	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
Discipline - Mechanical Calibration									
1	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Plunger Dial Gauge	Instrumental Error	0-1 mm (L.C.0.001 mm)	Plunger In / Out 0.12mm,0.30mm,0.48mm,0.64mm, 0.86mm,1.00mm	The CMC of participant laboratory should be coarser than 1 µm	Running Round	₹ 8500 ₹ 7000
2	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Analog Lever Dial Gauge	Instrumental Error	L.C. 0.01 mm Range 0 to 0.8 mm	Upward Direction (forward /Reverse) 0.10mm,0.50mm and 0.80 mm Downward Direction (forward /Reverse) 0.10mm,0.50mm and 0.80 mm	the CMC of participant laboratory should be coarser than 1.4 µm	Running Round	₹ 8500 ₹ 7000
3	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Bore Gauge With Dial	Instrumental Error	18-35 mm with dial 0.001 mm	0.02mm, 0.15mm,0.36mm,0.58mm, 0.82mm,1.00mm	The CMC of participant laboratory should be coarser than 1 µm	Running Round	₹ 8500 ₹ 7000
4	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Plain Plug Gauge	Outer Diameter	up to 100 mm	ø 5.00 mm ø36.50 mm ø 98.00 mm	The CMC of participant laboratory should be coarser than 1 µm	Running Round	₹ 8500 ₹ 7000
5	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Micrometer Setting Standard	Length	up to 100 mm	25 mm & 75 mm	The CMC of participant laboratory should be coarser than 1 µm	Running Round	₹ 8500 ₹ 7000
6	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Plain Taper Ring Gauge	Outer Diameter	up to 100 mm	25mm 35mm	The CMC of participant laboratory should be coarser than 1 µm	Running Round	₹ 8500 ₹ 7000
7	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Thread Ring Gauges	Simple Effective Diameter	up to 100 mm	M 5 X 0.8 -6g NOGO, M 40 X 3.0 -6g GO and G3 X11 ISO-228 B NOGO	The participant laboratory's CMC should be coarser than 1.1 µm	Running Round	₹ 8500 ₹ 7000
8	Dimension (Precision-Precision Instruments)	Linear – Reference (Measuring) Equipment	Tungsten Gauge Blocks	Central Length Deviation	up to 100 mm	1.5, 40 & 90 mm	The participant laboratory's CMC should be coarser than 0.02 + L/ 3000 µm L is in mm	Apr 26 to Jun 26	₹ 12000 ₹ 10000

SN	Categoery	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
9	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Digital Caliper	Instrumental Error	0-300 mm L.C. - 0.01 mm	Ext. Measurements 50mm, 150mm,250mm Int. Measurements 50mm, 150mm,250mm Depth Measurements 50mm	The CMC of participant lab. should be coarser than 11 µm	Running Round	₹ 8500 ₹ 7000
10	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Taper Plug Gauge	Diameter & Angle	Upto 50mm	∅ 50mm ∅ 38mm	The CMC of participant lab. should be coarser than 1.2	Running Round	₹ 8500 ₹ 7000
11	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Taper Thread Plug Gauge	Simple Effective Diameter	up to 100 mm	3" BSPT SY -B	The CMC of participating lab. should be coarser than 2 µm	Running Round	₹ 8500 ₹ 7000
12	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Thread Plug Gauge	Simple Effective Diameter	up to 100 mm	M8 X 1.25 6H Go M36 X 1.5 7H Go 2 X 12 UNC – 2B SPL NOGO	The CMC of participating lab. should be coarser than 1.3 µm	Running Round	₹ 8500 ₹ 7000
13	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Master Ring	Diameter	up to 100mm	5mm,26mm and 76mm	The participation Laboratory should have CMC coarser than 0.8 µm	Running Round	₹ 8500 ₹ 7000
14	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Granite Surface Plate (On Site)	Flatness	1000 X 630 mm	NA	The participating Laboratory should have CMC coarser than 3.5 µm	Open up to Mar 2026	₹ 8500 ₹ 7000
15	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Feeler Gauge	Thickness	up to 1 mm	Three Leaf	The participation Laboratory should have CMC coarser than 0.65 µm	Running Round	₹ 8500 ₹ 7000
16	Dimension (Basic-Measuring Instrument, Gauges etc.)	Gauges	Cylindrical Pins	External Diamteter	Up to 20 mm	Three Pins	The participation Laboratory should have CMC coarser than 0.65 µm	Running Round	₹ 8500 ₹ 7000
17	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Analog Durometer	Hardness Shore A and Shore D	0 - 100	Shore A - 0,10, 20,30, 40,50,60,70,80,90 and 100 Shore D - 0,10, 20,30, 40,50,60,70,80,90 and 100	The participation Laboratory should have estimated uncertainty coarser than 1.16% for Shore A and for Shore D 1.16 %	Running Round	₹ 10000 ₹ 8000 (Per Parameter) For both ₹ 12000.00

SN	Category	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
18	Dimension (Basic-Measuring Instrument, Gauges etc.)	Dimension	Digital External Micrometer	Instrumental Error	0- 25 mm L.C. 0.001 mm	2.5 , 10.3, 25 mm	The participation Laboratory should have CMC coarser than 0.90 µm	Running Round	₹ 8500 ₹ 7000
19	Dimension (Precision-Precision Instruments and Surface Topology)	Electronic Height Gauge	2D Digital Height Gauge	Instrumental Error	0-600 mm L.C. 0.0001 mm	50,100,200,300, 500 & 600 mm	The participation Laboratory should have CMC coarser than 2 µm	Running Round	₹10,000.00
20	Dimension (Basic-Measuring Instrument, Gauges etc.)	Length	Dial Thickness Gauge	Length	0-10mm L.C. 0.01mm	2.5 mm, 5.1mm and 7.7mm	The participation Laboratory should have CMC coarser than 4.2 µm	Running Round	₹ 7000
21	Dimension (Basic-Measuring Instrument, Gauges etc.)	Depth Micrometer	Depth Micrometer	Instrumental Error	0-25mm L.C. 0.001mm	4.8 mm, 17.3 mm and 25mm	The participation Laboratory should have CMC coarser than 4 µm	Running Round	₹ 7000
22	Dimension (Basic-Measuring Instrument, Gauges etc.)	Go-NOGO Snap Gauges	Snap Gauge	Gap Size	up to 100mm	81.6 mm Go and NOGO	The participation Laboratory should have CMC coarser than 0.8 µm	Running Round	₹ 7000
23	Pressure & Vacuum Gauge	Industrial Pressure Gauge (Ana/Dig)	Digital Pneumatic Pressure Gauge	Pressure	0 - 50 bar L.C.0.01 bar	5,10,15,20 and 35 bar	The CMC of participating lab. Should be coarser than 0.005% rdg	Running Round	₹ 11,500.00
24	Pressure & Vacuum Gauge	Industrial Pressure Gauge (Ana/Dig)	Hydraulic Pressure Gauge	Pressure	0-700 bar L.C. 0.01 bar	100,200,350,500 and 650 bar	The CMC of participating lab. Should be coarser than 0.009% rdg	Running Round	₹ 14000 ₹ 11500
25	Pressure & Vacuum Gauge	Industrial Pressure Gauge (Ana/Dig)	Hydraulic Pressure Gauge (Low)	Pressure	0-200 bar L.C. 0.01 bar	50,100,125,150,175 bar	The CMC of participating lab. Should be coarser than 0.009% rdg	Running Round	₹ 14000 ₹ 11500
26	Pressure & Vacuum Gauge	Industrial Pressure Gauge (Ana/Dig)	Digital Vacuum Gauge	Pressure	-1 to 0 bar L.C. 0.01 bar	-0.1,-0.3,-0.5,-0.7 and -0.8 bar	The CMC of participating lab. Should be coarser than 0.009% rdg	Running Round	₹ 14000 ₹ 11500
27	Pressure & Vacuum Gauge	Industrial Pressure Gauge (Ana/Dig)	Analog Hydraulic Pressure Gauge	Pressure	0-600 bar L.C. 5 bar	50,200,300,400 &550 bar	The CMC of participating lab. should be coarser than 0.009% rdg	Running Round	₹ 14000 ₹ 11500

SN	Categoery	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
28	Pressure & Vacuum Gauge	Industrial Pressure Gauge (Ana/Dig)	Absolute Pressure Gauge	Pressure	0- 4 bar L.C. 0.0001 bar Absolute	0.4 bar, 0.8 bar, 1.2 bar , 3 bar , 4 bar	The participation Laboratory should have CMC coarser than 0.003% rdg (abs)	Running Round	₹ -14000 ₹ 12500
29	Force	Force	Load Cell	Force	upto 500 N L.C.0.01 N	50,100,150,200,250,300,350,400, 450 and 500 N	The participant laboratory's CMC should be coarser than 0.01% rdg	Open up to Mar 2026	₹ 14000 ₹ 11500
30	Force	Force	Push Pull Gauge (Force Gauge)	Force	up to 1000 N L.C. 0.1 N	100,200,500,700 and 1000 N	The CMC of participating lab. should be coarser than 0.07%	Open up to Mar 2026	₹ -14000 ₹ 11500
31	Torque Generating & Measuring devices	Torque Transducer	Torque Sensor	Torque	up to 1000Nm L.C. 0.1Nm	25,100,150,175,500, 700 & 1000 Nm	The CMC of participating lab. should be coarser than 0.01%	Open up to Mar 2026	₹ -36000 ₹ 25000
32	Torque Generating & Measuring devices	Torque Wrenches (Type I & II)	Digital Torque Wrench	Torque	Range 1-500 Nm L.C.0.1 Nm	50,300, 500Nm	The CMC of participating lab. should be coarser than 0.3%	Open up to Mar 2026	₹ -15000 ₹ 12000
33	Mass and Volume	Mass	Digital Weighing Balance (On site)	Mass	220 g L.C. 0.1 mg 3000g L.C. 1 mg Using E1 &E2 Class weights	For 220 g - 50 g,100g and 200 g- Participant can use only E1 Class Weights For 3000g - 50g,100g,200g,500g, 1000g,1500g and 2000g- Participants can use E1/E2 Class Weights	The participation Laboratory should have CMC coarser than 0.025 mg for 220 g & 42 mg for 3000g	Open up to Mar 2026	₹ -14700 ₹ 11500 (per balance)
34	Mass and Volume	Mass	F1 Class Weights	Mass	up to 500 g	Mass 2g (E1 Class), 20g (E1 class), 200g (E1 Class) and 500g (E2 Class) Participant has to use E1/E2 Class Weights	The participant laboratory's CMC should be coarser than 0.002 mg for 2 g, 0.005 mg for 20g, 0.026 mg for 200 g and 0.25 mg for 500 g	Open up to Mar 2026	₹ -14000 ₹ 11500
35	Mass and Volume	Mass	Digital Weighing Balance Class III Resolution 1000 mg	Mass	0-15 Kg	For 0.01Kg , 0.05Kg, 0.5Kg, 1Kg,2Kg, 5Kg,7Kg,10Kg,12Kg,15Kg - Participant can use only F1 or better Class Weights	The participation Laboratory should have for CMC coarser to 650 mg	Open up to Mar 2026	₹ -14700 ₹ 11500
36	Acoustics	Sound Level Meters	Sound Level Meter	Sound Level	94dBA	94dBA @1KHz	The participation Laboratory should have CMC coarser than 0.8 dBA	Open up to Mar 2026	₹ 7000
Discipline - Electro-Technical Calibration									

SN	Categoery	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
37	Alternating Current (<1GHz) Direct Current	High Voltage	HV Probe / Divider with Digital Multimeter	AC / DC High Voltage	DC V -1kV- 40kV AC V- 1kV-28kV 50Hz	AC- 1,5,10,15 and 25 kV 50 Hz DC -1,10,20,30 and 40 kV	The participant laboratory's CMC should be coarser than 1.4% for AC & DC High Voltage	Running Round	₹ 15000
38	Frequency and Resistance	Frequency AC/ DC Current	DC Resistance & Frequency (41/2 Resolution)	Frequency DC Resistance	DC Resistance Frequency	10Ω, 1kΩ, 10kΩ, 1 MΩ,10 MΩ 50Hz, 1kHz,10kHz,50kHz, 100kHz	The participating Laboratory should have CMCs coarser than 0.0003 % to 0.025 % for DC Resistance 0.0002% for Frequency.	Running Round	₹ 20000 ₹ 8500 (Per Parameter) For both parameters ₹ 15000.00
39	Temperature Simulation	Temperature Simulation	Process Calibrator - Source Mode	Temperature Simulation	Type K , PT 100 Type S, Type J	Type K- 35, 275, 515,790 and 1065°C RTD PT 100 (-50),180,410,590 and 770 °C Type S - 50,370,690,1010 and 1330 in °C Type J- 35, 275, 515,790 and 1065°C	The participation Laboratory should have uncertainty coarser than 0.15 °C for K Type and 0.09 °C for RTD PT 100 Sensor 0.12 °C for J Type and 0.48 °C for s Type	Running Round	₹ 20000 ₹ 10000 for any two parameters for four parameter ₹ 15000
40	Alternating Current (<1GHz) Alternating Voltage	Current Voltage	Digital Multimeter 3 1/2 Resolution Make: Fluke Source Mode	Current DC & DC Voltage Current AC & AC Voltage 50Hz	DC Voltage: 50mV to 980V DC Current: 100µA to 10A AC Voltage:50mv to 980V AC Current: 100µA to 10A	DC Voltage - 2mV,200mV,2V,200V, 1000V DC Current - 10mA, 100mA, 1A,5A, 10A AC Voltage - 100mV, 200mV,2V,200V,700V at Frequency 50Hz AC Current - 10mA, 100mA,1A,5A, 10A at Frequency 50Hz	The participation Laboratory should have uncertainty coarser than 3.04% to 0.029% for DC Voltage 0.128% to 0.058% for DC Current 0.083% to 0.06% for AC Voltage 0.182% to 0.078% for AC Current	Running Round	₹ 20000 ₹ 8500 (Per Parameter) For both AC& DC parameters ₹ 15000.00

SN	Categoery	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
41	Alternating Current (<1GHz) Alternating Voltage	Current Voltage	Digital Multimeter 4 1/2 Resolution Make:GW Instek Source Mode	Current AC & AC Voltage 50Hz Current DC and DC Voltage	Range & Resolution - Voltage - 220.00mV, 2.2000V, 22.000V, 220.00V, 1000.0V Current - 220.00µA, 2200.0µA, 22.000mA, 220.00mA, 20.000A	AC Voltage- 100 mV, 200 mV, 2V, 200V, 700 V at Frequency 50Hz DC Voltage- 2mV, 200 mV, 2 V, 200 V, 1000 V AC Current - 10mA, 100mA, 1A, 5A,10A at Frequency 50Hz DC Current - 10mA, 100mA, 1A, 5A,10A	The participation Laboratory should have uncertainty coarser than 0.006 % to 0.023 % for DC Voltage 0.046% to 0.056% for DC Current 0.016% to 0.038% for AC Voltage 0.054% to 0.156% for AC Current	Running Round	₹ 20000 ₹ 8500 (Per Parameter) For both AC& DC parameters ₹ 15000.00
42	Alternating Current (<1GHz) Direct Current	AC Resistance (at 1Kz) Capacitance (at 1Kz) Inductance (at 1Kz)	LCR Meter	AC Resistance (at 1Kz) Capacitance (at 1Kz) Inductance (at 1Kz)	(at 1KHz) AC Resistance – 1 Ω to 10 K Ω Capacitance -1nF – 10µF Inductance - 100µH- 1H	AC Resitance 1 Ω, 10 Ω, 100 Ω,1K Ω& 10K Ω 1KHz Capacitance - 1nF,10nF,100nF,1µF 1KHz Inductance - 1mH, 10mH, 100mH,1H 1 KHz	The participation lab should have uncertainty coarser than For Range 100 uH to 1 H – Inductance 0.09% to 0.386% For Range 1nF to 1 uF – Capacitance 0.061% For Range 1 Ω to 10KΩ – AC Resistance 0.07%	Running Round	₹ 10000 ₹ 8500
Discipline - Thermal Calibration									
43	Specific Heat & Humidity	Humidity	Thermo Hygrometer	Humidity & Temperature	Rh 5% to 95% at 25°C to 50°C L.C.0.01% & Temp up to 50°C L.C. 0.01°C	Humidity 20%Rh, 30%Rh,50%Rh, 70%Rh & 85%Rh @ 25°C & 10°C, 20°C, 30°C,40°C to 50°C @50%RH	The participant laboratory's CMC should be coarser than 0.45 %Rh for Humidity and 0.15°C for specific Heat Temp	Running Round	INR 15000 (For both Parameters) INR 8500 (One parameter)
44	Specific Heat & Humidity	Humidity	Thermo Hygrometer	Humidity	5 to 50 °C, 20 to 95%Rh Humidity L.C. 0.1°C	Humidity 30%Rh, 40%Rh,50%Rh, 65%Rh & 80%Rh @ 25°C	The participant laboratory's CMC should be coarser than 0.45%Rh	Open up to Mar 2026	₹ 10000 ₹ 8500
45	Radiation Pyrometer	Temperature	IR Sensor	Temperature	upto 1200°C L.C. 0.1 °C	100°C,200°C, 350°C,500°C,700°C,800°C,950°C, 1000°C & 1200°C,	The participant laboratory's CMC should be coarser than 3.44 °C	Open up to Mar 2026	₹ 14000 ₹ 8500

SN	Categoery	Subcategory	PT Item Name	Parameter	Range	Measurement Points	Participation Criteria	Round Schedule	PT Charges
46	Temperatures Indicator with Sensor	Temperature	RTD PT 100 Sensor with Display 0.01 °C	Temperature	50°C to 250°C L.C.0.01°C	50°C, 100°C, 150°C,200°C & 250°C	The Participant Laboratory's CMC should be coarser than 0.1°C	Running Round	₹ 12000 ₹ 8500
47	Temperatures Indicator with Sensor	Temperature	RTD PT 100 Sensor with Display 0.01 °C	Temperature	-50°C to 250°C L.C.0.01°C	-50°C, 0°C, 75°C, 125°C, 175°C, & 225°C,	The Participant Laboratory's CMC should be coarser than 0.1°C	Running Round	₹ 12000 ₹ 8500
48	Thremo couple (Without indicator)	Temperature	S Type Thermocouple	Temperature	S Type 400°C to 900°C	400°C, 500°C, 650°C,800°C and 900°C	the participant laboratory should have CMC coarser than 1.5 °C	Running Round	₹ 12000 ₹ 8500
For online registration please visit our website OMS.OCTAGON.CO.IN									
	Every new PT round will start with minimum 5 PT registration for that PT round ; We consider and allow minimum 5- 10 participants in one PT round								
	GST / IGST as applicable 18% extra will be applicable, in addition to the above mentioned charges.								

Rev.01 DT: 24/12/2025